

BVDV Positive Control

Product No.: 1506736

(For Research Use Only)

■ Product description

SHENTEK® BVDV Positive Control is intended for use with SHENTEK® Virus DNA & RNA Extraction and Detection Kits, and serves as a positive control to monitor the entire process of BVDV RNA extraction and RT-PCR detection assay.

The product described herein is for research use only and is not intended for diagnostic or therapeutic use.

■ Kit components and storage

Reagent	Quantity	Storage
BVDV Positive Control	100 µL × 3 tubes	below -65°C

BVDV Positive Control can be stored under the appropriate conditions for up to 12 months. Please check the expiration date on the labels.

■ Attentions

1. Thaw one tube completely at 2-8°C or on ice, then vortex and spin briefly.
2. Follow the Positive Control Sample (PCS) preparation protocol outlined in the user guide of SHENTEK® BVDV Detection Kit and SHENTEK® Bovine Virus Detection Kit (BVDV/REO-3/PI-3/BPV-3/BAV-3).
3. This product is ready for direct use without dilution.
4. Use only after it is equilibrated to room temperature, and avoid repeated freeze-thaw cycles.
5. Do not mix reagents from different batches.
6. Do not use reagents beyond their expiration date.
7. As an armored RNA product, this material should be handled in a biosafety cabinet. Experimental waste generated during its use must be sterilized via autoclaving and disposed of in compliance with medical waste regulations.

■ Related products

Product Name	Product No.	Quantity
SHENTEK [®] Virus DNA & RNA Extraction Kit	1506730	50 Extractions
SHENTEK [®] BVDV Detection Kit	1506731	50 Reactions
SHENTEK [®] Bovine Virus Detection Kit (BVDV/REO-3/PI-3/BPV-3/BAV-3)	1506741	50 Reactions

Effective date: 12 Mar. 2026

Support & Contact**SHENTEK**

Huzhou Shenke Biotechnology Co., Ltd.

www.shentekbio.com

Address: 8th Floor, 6B Building, No.1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China

E-mail: info@shentekbio.com

Phone: +1 (908) 822-3199 / (+86) 400-878-2189